

October 19, 2017

**ADDENDUM No. 2**

**Request for Proposals CS-17-10  
SUPPLY AND DELIVERY OF A  
PORTABLE SPECTRUM ANALYZER**

*The addendum is being issued prior to the closing of the Request for Proposals (RFP) to provide further information, make changes to, or to clarify the RFP Documents and is to be read, interpreted and coordinated with all other parts of the RFP Documents. In the case of a conflict with the balance of the documents, this Addendum shall govern. **Proponents shall attach a signed copy of this addendum to their tender submission, failure to do so may result in a non-compliant bid.** This addendum shall form part of the Contract Documents.*

**1. QUESTION:**

The RFP specifies 6 GHz as the max frequency for the directional antennas, is this the maximum frequency you need to cover?

**ANSWER:**

The frequency range specification for the analyzer is 9KHz to 6 GHz.

**2. QUESTION**

The DANL requirement isn't clear to me. Can you clarify?

**ANSWER:**

The DANL specification for the analyzer is <-167 dBm/Hz, **the specification of "<7 dB for frequency's less <3GHz" has been removed.**

I/We hereby verify that we have considered this addendum in our bid submission.

\_\_\_\_\_  
Bidder Signature

\_\_\_\_\_  
Date

All inquiries relating to this ITT must be directed to:  
Tony Fry, Public Safety Coordinator, Project Manager  
Email: [tfry@rdffg.bc.ca](mailto:tfry@rdffg.bc.ca)